

FidoMasks

FidoMasks

. - MASKS



3-Piece Half Mask APPROVAL LABEL



FIDO MASKS CO., LTD.

No. 31, Gong 1st Rd., Dajia Dist., Taichung City 437, Taiwan (R.O.C.) TEL: +886-4-2682-2737

THIS RESPIRATOR IS APPROVED ONLY IN THE FOLLOWING CONFIGURATION:

| TC- | Protection ¹ | Respirator F810CI | | | Cautions and Limitations ² |
|----------|-------------------------|----------------------|----------|-------------|---------------------------------------|
| 10 la. | | | | | 125 |
| 0,0 | | Filter Frame | Filter | Filter Lock | Wya |
| - | | RFF | F810CI-F | RFL | · 701. |
| 84A-8087 | N95 | X | X | X | ABCJMNOP |

1. Protection

N95 - Particulate Filter (95% filter efficiency level) Effective against particulate aerosols free of oil; time use restrictions may apply

2. Cautions and Limitations

- A Not for use in atmospheres containing less than 19.5% oxygen.
- B Not for use in atmospheres immediately dangerous to life or health.
- C Do not exceed maximum use concentrations established by regulatory standards.
- J Failure to properly use and maintain this product could result in injury or death.
- M All approved respirators shall be selected, fitted, used, and maintained in accordance with MSHA, OSHA and other applicable regulations.
- N Never substitute, modify, add, or omit parts. Use only exact replacement parts in the configuration as specified by the manufacturer.
- O Refer to User's Instructions, and/or maintenance manuals for information on use and maintenance of these respirators.
- P NIOSH does not evaluate respirators for use as surgical masks.

FidoMasks

ridaMasks